

DOC-SS004 DECLARATION OF COMPLIANCE
 Slot Powder Lance

Trade names: SteriWare® Slot Lance (Multi level powder sampler)

D1050 White non-sterile
 Overall length 970mm
 Nominal sample size 170ml
 Overall diameter 25mm
 Material HDPE
 20 pieces per box




Supplier/Importer: Klipspringer Ltd
Address: Rynor House, Farthing Road, Ipswich, Suffolk, UK. IP1 5AP
Telephone: 0044 (1) 461800

- We confirm that the goods are manufactured with Polypropylene that conforms to:
 - FDA CFR 177.1520
 - EU regulation 10/2011 (Plastic items intended for contact with food)
 - EC regulation 1935/2004 (Items intended for contact with food).
- An FDA acceptable blue dye (less than 1%) is added during manufacture to make the scoops blue.
- The HDPE resin used in the manufacture of the above products is BSE/TSE free. During manufacture this item does not come into contact with any products or processes that may pose a BSE/TSE risk.
- These goods are moulded and packed in a class 100,000 cleanroom. They are individually packed.
- This product is designed for single use only. It must not be used multiple times as it may pose a foreign body risk.

This document was prepared on behalf of Klipspringer Ltd and the information included is to the best of our knowledge correct at the time of writing. Klipspringer offers the information within this document as a guide only, they do not represent any guarantee of the prescribed products in the sense of the legal guarantee regulations. It is the responsibility of the end user to ensure the items purchased are suitable for the intended application.

Supplier/Importer: Klipspringer Ltd
Address: Rynor House, Farthing Road, Ipswich, Suffolk, UK. IP1 5AP
Telephone: 0044 (0) 1473 461800
Email/Website: info@klipspringer.com / www.klipspringer.com

Sheena Britton
 Technical Compliance Manager
 Klipspringer Ltd
15-07-2016



Date of issue	15-07-2016	Revision No.	001	Revised by	SB
Authorised by	S. Britton	Document No.	DOC-SS004	Page 1 of 1	